

STAND-AID SLING



Product Code: HSL500
HSL510 (With Clips)

- Description:**
- The Stand-Aid Sling is designed to assist with the standing process using a hoist
 - It is fully padded with non-slip mesh and adjustable body belt for added security
 - Only suitable for those who have a degree of weight bearing capability
 - Comes with Velcro pads to protect the sling from damage by contact with the velcro
 - Woven label to increase the life of the sling
 - UK manufactured
 - Extra stitching to increase the quality and safe working load
 - Available in a range of sizes
 - Secured by loop or clip-fastening
 - The HSL520 is specially designed to fit with Harvest Healthcare's Ascent Stand-Aid

Safe Working Load: 400kg / 62 Stone (Polyester)

Cleaning Instructions: Each sling can be laundered up to 80°C using a mild soap solution and then rinse. Slings will air dry quickly in a warm room or in the normal manner. Tumble dry on low setting only. Do not place over a convector heater or steam pipes. Do not use Biological detergents. Do not use Bleach. Do not use strong chemicals. A faded label is an early indicator that sling is subject to harsh / abrasive cleaning methods do not continue to use sling. All slings fitted with a Velcro body belt should have the Velcro square removed during laundering and refitted afterwards. This will significantly extend the lifespan of the product.

Supplier Contact Details:

t: +44(0)1709 377172 e: enquiries@harvesthealthcare.co.uk w: www.harvesthealthcare.co.uk
Sheaf house, Bradmarsh Way, Bradmarsh Business Park, Rotherham UK, S60

Important Safety Notice

Harvest Healthcare Ltd supplies equipment and appliances to cover a range of medical conditions. Specific choice of equipment is the sole responsibility of the clinical practitioner or carer overseeing individual patient care. Harvest Healthcare Ltd supplies a range of products such as foam mattresses, active air mattresses, fall protection mats and bed rail protection bumpers. These products are usually covered in vapour permeable PU fabric that could cause risk of suffocation in the event of prolonged facial contact.

(Mattress risk categories are to be used as a guide only)

Risk assessments specific to user must be carried out prior to the use of all products. Products and data continually undergo development and may therefore deviate from the details given.